Review

Plasmonic and Hybrid Whispering Gallery Mode–Based Biosensors: Literature Review

Maurizio Manzo¹, PhD; Omar Cavazos¹, MSc; Zhenhua Huang², PhD; Liping Cai², PhD

¹Photonics Micro-Devices Fabrication Lab, Department of Mechanical Engineering, University of North Texas, Denton, TX, United States

²Department of Mechanical Engineering, University of North Texas, Denton, TX, United States

Corresponding Author:

Maurizio Manzo, PhD

Photonics Micro-Devices Fabrication Lab

Department of Mechanical Engineering

University of North Texas

3940 N Elm St, F115

Denton, TX, 76207

United States

Phone: 1 940 369 8266

Email: maurizio.manzo@unt.edu

Abstract

Background: The term “plasmonic” describes the relationship between electromagnetic fields and metallic nanostructures. Plasmon-based sensors have been used innovatively to accomplish different biomedical tasks, including detection of cancer. Plasmonic sensors also have been used in biochip applications and biosensors and have the potential to be implemented as implantable point-of-care devices. Many devices and methods discussed in the literature are based on surface plasmon resonance (SPR) and localized SPR (LSPR). However, the mathematical background can be overwhelming for researchers at times.

Objective: This review article discusses the theory of SPR, simplifying the underlying physics and bypassing many equations of SPR and LSPR. Moreover, we introduce and discuss the hybrid whispering gallery mode (WGM) sensing theory and its applications.

Methods: A literature search in ScienceDirect was performed using keywords such as “surface plasmon resonance,” “localized plasmon resonance,” and “whispering gallery mode/plasmonic.” The search results retrieved many articles, among which we selected only those that presented a simple explanation of the SPR phenomena with prominent biomedical examples.

Results: SPR, LSPR, tilted fiber Bragg grating, and hybrid WGM phenomena were explained and examples on biosensing applications were provided.

Conclusions: This minireview presents an overview of biosensor applications in the field of biomedicine and is intended for researchers interested in starting to work in this field. The review presents the fundamental notions of plasmonic sensors and hybrid WGM sensors, thereby allowing one to get familiar with the terminology and underlying complex formulations of linear and nonlinear optics.

(*JMIR Biomed Eng* 2021;6(2):e17781) doi: [10.2196/17781](https://doi.org/10.2196/17781)

KEYWORDS

plasmonic; whispering gallery mode, microlasers, biomedical; sensors

Introduction

The term “plasmonic” describes the relationship between electromagnetic fields and metallic nanostructures [1]. Plasmonic sensors have attracted great interest from researchers and engineers alike. Surface plasmon resonances (SPRs) are electromagnetic waves that are produced when a metal nanostructure (ie, spherical or cylindrical) interacts with a

dielectric material [2,3]. The interesting optical characteristics of surface plasmons have made many important contributions to the field of medicine [4]. For example, highly sensitive plasmonic sensors have been developed to detect many kinds of cancers [5], and based on the SPR concept, a plasmonic interferometer array–based sensor was developed for detecting cancers [6].

This paper discusses the physical principles in brief and introduces several methods employing plasmonic systems such as SPR and localized plasmon resonance. SPR methods have attracted great interest in biomedical applications. This technique entails observing small changes in the refractive index of the combination of dielectric materials and metal [7]. In addition, plasmonic nanoparticles and nanostructures have been used in biosensing applications. These structures are typically made of noble metals such as gold and silver [8,9]. The cytotoxicity of these metals based on their concentrations are under investigation and studies have shown potential biomedical applications for these metals at certain concentrations [10,11]. These nanostructures are used in photoacoustic imaging and phototherapy. For example, gold nanorods with varied light absorption peaks have been used in imaging and theranostics [12]. Plasmonic sensors also have been used in biochip applications and biosensors [13-15]. Other techniques such as localized plasmon resonance have also been utilized in various biomedical applications [16]. In summary, plasmon-based sensing methods are indispensable tools for sensing in the field of biomedicine. Moreover, these devices have the potential to be implemented as implantable point-of-care devices [17-19].

Methods

We performed a literature search on ScienceDirect for studies on plasmonic and hybrid whispering gallery mode (WGM) sensors and retrieved more than 3400 articles (both research and review articles) published in the field of medicine and dentistry. According to their characteristics, sensors were grouped into 4 categories, namely, SPR, localized SPR (LSPR), tilted fiber Bragg grating (TFBG), and hybrid WGM sensors.

The search for articles related to SPR and LSPR was straightforward. The search results retrieved many articles, among which we selected only those that presented a simple explanation of the aforesaid phenomenon with prominent biomedical examples. Within these 2 fields, a third field was categorized (TFBG) due to the prominent presence of biosensor-based devices that utilize the TFBG principle.

The keyword “whispering gallery mode” was associated with the term “plasmon” and only retrieved 3 papers in the pharmacology, toxicology, and pharmaceutical fields. Therefore, for the latest category (ie, hybrid WGM sensors), the search was expanded to the engineering field and eventually 35 reports were identified. More papers were found in other fields such as physics, astronomy, material science, and chemistry.

Results

Many papers, for example [20-33], describe the physical principle of SPR, and provide the definitions and discuss exemplary applications to illustrate how changing the refractive index can be used for sensing through the plasmonic effect and how the light is generally coupled to the biosensor. A total of 6 papers [23-27,29] illustrated that optical fibers can be used

in conjunction with SPR for sensing applications. Also, 13 papers were selected to discuss SPR-based metal nanostructures [2,20-22,34-42].

LSPR is discussed based on 11 papers [36-39,43-49]. As before, the physical principle and 2 representative examples are provided to understand the main differences between LSPR and the previous methods.

Although TFBG could be associated with the SPR-based optical fiber sensing method, many different papers have been found on this topic, and therefore, a separate category (ie, tilted fiber Bragg grating) was created. Several papers are used to illustrate the physical principle and applications.

Lastly, 20 papers were used to introduce the WGM and the hybrid WGM sensing [34,35,40-42,50-63]. This type of sensing methodology was not directly related to the medical literature, but biomedical applications were proposed and the future implementation of this method is likely to become the gold standard in some areas. This review paper presents and discusses the sensing techniques, including SPR, LSPR, TFBG, and hybrid WGM, as well as their applications using representative examples in the biomedical fields.

Discussion

Surface Plasmon Resonance

The term “plasmonic” describes the relationship between metallic structures and dielectrics in an electric field. The oscillations of electrons between a metal sample and a dielectric field are referred to as SPR. The attenuated total reflection (ATR) configuration is one of the prism coupling-based SPR methods [20-22]. In ATR, a metal sheet is placed on top of a light coupling substrate, such as glass (Figure 1). The light source is then directed into a prism and a detector gathers the resonances. Therefore, the resonance is displaced as a sharp dip in the output spectrum due to the absorption of the surface plasmon wave [20]. In a previous study [20], the ATR configuration was used to monitor the refractive index of the human skin as shown in Figure 1. Besides, this same configuration has been used for monitoring humidity, where the effect of the temperature on the sensor was analyzed. The sensor consisted of a chalcogenide substrate layer, gold layer, and buffer layer [22].

Optical fiber sensing with SPR has been used to detect different kinds of biological targets such as antibodies. The phenomenon of SPR occurs on the surface of the optical fiber [23]. Optical fiber SPR methods have advantages over traditional prism methods such as the ATR technique, which can be explained as follows: optical sensors use remote sensing and optical fibers have a reasonably lower cost and a more compact size. In addition, these types of sensors provide label-free sensing with high sensitivity [24]. In some cases, finite-element methods were used to analyze the design of the optical fiber sensors [25,26].

Figure 1. Proposed SPR sensor probe setup for the determination of refractive index of human skin tissues. IB: incoming light beam; SB: sensed light beam. Reproduced, with permission, from Elsevier [20].

Another class of optical fiber based on SPR uses photonic crystal fibers, which have been used as alternatives to traditional optical fibers. A notable advantage of these fibers is that they have more controllable birefringence and therefore a better control on light propagation and confinement directions [27]. Photonic crystals are dielectric materials that have a periodicity (repeated optical structure) in two or three dimensions. They are usually fabricated by etching, which can form a photonic bandgap, allowing to configure the light for different uses. The bandgap depends on the structural content of the crystal, such as refractive index and periodicity. The peak frequency shown in the transmission spectrum depends on the shape and size of the lattice defects [28].

Another method for SPR is the use of metallic nanostructures. Gold and silver nanostructures have been extensively used in past years, because they can be characterized by their size and shape [29]. Silver nanorods are suitable for biomolecular detection [30]. In addition, the iron oxide–gold nanoparticles can experience both plasmonic and magnetic phenomena, which allows for their use in different biomedical applications. One advantage of this type of particle is that it can be moved due to its magnetic property and still demonstrate plasmonic behavior. This can tremendously facilitate the analysis of biological targets

[31]. In another study, a piece of portable SPR instrument was developed using nanoparticles, which was able to detect testosterone [32]. Another type of SPR sensor is the plasmonic waveguide. Plasmonic waveguide designs tend to be suitable for chipping applications. This is mainly because of their compactness and the use of SPRs [33].

Localized Surface Plasmon Resonance

Localized SPR (LSPR) is the amplitude of oscillation of free electrons that occurs at a certain frequency, which can be used to detect biomolecules such as proteins in real time. In one study, a gold nanoplasmonic sensor was used to detect cancer markers in clinical samples. The sensor could also detect proteins such as biotin (Figure 2). Besides, it has the potential to detect DNA [43]. LSPR is mainly related to nanostructure/nanoparticles such as nanorods. In addition, LSPR does not require coupling, for example, with prisms and is easy to operate. Therefore, this method is widely used in the scientific field [44]. The type of nanostructure used has an impact on the strength of LSPR. For example, nanostars can be used for tuning the sensors and promoting a strong LSPR signal. However, structures such as nanorods and nanospheres have widely been investigated for various imaging applications [45].

Figure 2. Detection of protein-protein binding event on the gold nanostructure through LSPR peak shift. Licensed under Creative Commons Attribution 4.0 by the authors [43]. LSPR: localized surface plasmon resonance.

Gold nanorods have been used as plasmonic sensors for detecting mercury. The deposition of mercury on the nanorods was observed by monitoring the LSPR shifts using darkfield microscopy [46]. Tao et al [47] used a gold and silver alloy nanoplasmonic device to detect mercury concentrations. LSPR has also been used to develop silk plasmonic absorber sensors [36]. In this case, silk protein is used as an insulator in the insulator-metal resonator configuration. Besides, the silk plasmonic absorber sensor was applied as a glucose sensor, which demonstrated a high sensitivity of 1200 nm/RIU (refractive index unit) and high relative intensity change [36]. Metal nanoparticles have also been used for copper detection in samples, mainly because LSPR is influenced by the morphology and size of particles. Ding et al [37] used gold

nanoparticles to detect specific copper ions. Besides, LSPR sensors have been integrated into optical fiber devices. Tu et al [38] used hollow gold nanocages for LSPR optical fiber sensors. The sensor had a sensitivity of around 1933 nm/RIU. Furthermore, the sensitivity can be adjusted by changing the aspect ratio of the gold nanocages. Yousuf et al [39] developed a metal-insulator-metal configuration, which consisted of an elliptical nanorod, rectangular nanoslabs, and a metallic grating. Unser et al [48] developed a selective collagen gold nanoparticle-based sensor, which works based on the plasmonic coupling of the nanoparticles and the collagen fibrils. A redshift (toward the right side of the spectrum) in the LSPR frequency indicates the detection of glucose. Overall, the conjugates were able to detect glucose and heparin (Figure 3).

Figure 3. The 2 sensing schemes addressed in this work using collagen-nanoparticle conjugates. (A) The native collagen is added before the gold nanoparticles (AuNPs); (B) the collagen after it has been coated in AuNPs forming a collagen-nanoparticle scaffold; (C) in order to carry out biosensing measurements of glucose, the collagen nanoparticle scaffold is crosslinked by glucose after it has been incubated at 35°C and the covalent product glucosepane has formed; (D) lastly, the binding interactions between the collagen-nanoparticle scaffold and the heparin-coated 80-nm gold nanoparticles are used to detect heparin. Licensed under Creative Commons Attribution 4.0 by the authors [48].

Tilted Fiber Bragg Grating

In the configuration of TFBG, the refractive index modulation planes are in a tilted position, which helps to measure very small changes. These small changes can be fully analyzed by observing the refractive index of the fiber. The tilted grating disrupts the fiber's symmetry, which causes some core-guided lights to be coupled that allows the cladding mode resonances to be observed. These resonances are observed as a comb of sharp dips. It has been noted that these methods can greatly increase quality factors (ie, Q values of up to 10^4) [64].

Therefore, this type of sensing method can be widely used for biomedical applications in constricted spaces. Fiber optic-based sensors allow easy sensor installation. In one study, TFBG was

used to detect the variation in protein in the urine of rats [65]. Results were obtained using a coated TFBG embedded inside a microfluidic channel. The experiment was able to distinguish different kinds of urine. Results demonstrated a clear relationship between protein outflow and changes in the refractive index of the urine. This approach showed improvements in the detection of proteins at low concentrations [65]. The TFBG SPR sensor has been used for the detection of glycoprotein. Zhang et al [66] coated 10° TFBG with a 50-nm gold film to stimulate SPR on a sensor surface as shown in Figure 4. The sensor was able to distinguish between nonglycoproteins and glycoproteins. The TFBG-based sensor was also used to detect S-adenosyl-l-homocysteine (AdoHcy), with concentrations of up to 1 nM detected [67].

Figure 4. (A) Schematic of the tilted fiber Bragg grating (TFBG)-based surface plasmon resonance (SPR) (TFBG-SPR) sensor. (B) Transmission spectra of the sensor under P and S polarization. (C) Experimental setup. Licensed under Creative Commons Attribution 4.0 by the authors [66].

Hybrid Whispering Gallery Mode Sensors

WGM resonators have been used for different applications, especially for high-sensitivity and resolution sensors [40,50-55]. The WGM microstructure can be made in various types of shapes such as spheres, cylinders, and toroid [40,50-55]. The WGMs of the resonator can be observed by coupling light to the resonator. These electromagnetic waves circulate near the internal edge of the resonator [54]. Therefore, the resonances are generated by the total internal reflection of the confined light and when the optical path of the light is a multiple integer of the wavelength [55]. The WGM shift caused by an excitation method can be used to determine the change in the measured quantities [40,56]. The WGMs can be tuned by excitation

sources such as uniaxial stress and electric field. The theory is that the WGMs propagate across the pole of the spheres. Then, the deformation along the ends of the sphere and index of refraction change modify the position of WGM resonances [56,57]. WGM resonators have been developed using fused silica. Fused silica resonators have a quality factor (Q) of 10^9 [58]. Silica resonators have been used, but have low sensitivity because they exhibit high Young modulus and therefore high resistance to any deformation. Different materials such as polydimethylsiloxanes have been used to address the issue of low sensitivity [59]. The WGM resonances can be observed using different techniques. For example, a study used a charge-coupled device camera and a spectrometer to observe

the resonances, and therefore when doping WGM resonators with a laser dye material, the hybrid functions as tiny lasers that emit light under proper excitation conditions [40,50-55,67]. The light emitted from the resonator could then be coupled into the spectrometer using an optical lens setup [40]. Another study discussed a novel fiber-taper coupling system that couples light into microresonators. It has been observed that tapered optical fibers promote high coupling efficiency to the resonators. The experiment was completed with a silica microresonator coupled to a tapered optical fiber [34]. In a similar study, a silica microsphere resonator was critically coupled to a fiber taper. The fiber taper is useful because it allows for simple focusing and alignment of the input beam but uses the resonator as a passive element, which limits the application due to the presence of optical cabling [35,50].

Figure 5. Geometrical scheme of whispering gallery mode (WGM) microtoroid with a gold triangular nanoprism bound to its surface. Reproduced from Nadgaran H, Afkhami Garaei M. Enhancement of a whispering gallery mode microtoroid resonator by plasmonic triangular gold nanoprism for label-free biosensor applications. *Journal of Applied Physics* 2015 Jul 28;118(4):043101. [doi:10.1063/1.4927266], with the permission of AIP Publishing [42].

In other studies, polymeric WGM-based spherical resonators have been doped with metal nanoparticles to lower the energy required to activate the sensor [40]. In this case, the plasmonic effect enhanced light emission and lowered the energy threshold required for the structure to lase with higher temporal duration and more stable amplitude of the optical resonances, enabling multiplexed capabilities [40].

Electrically controlled graphene has also been applied to improve the performance of a hybrid silver-silica microdisk resonator. Most notably, the Q factor (energy stored and energy loss ratio) was improved and had a sensitivity higher than 1000 nm/RIU. Therefore, the hybrid sensor has a huge potential for use as a refractometer [60]. In recent years, there has been a growing interest in utilizing hybridization whispering gallery microstructures with the plasmonic effect. The motivation for this hybrid concept is that the single plasmonic sensing generates low Q factors (higher losses), whereas the presence of a resonant structure overcomes this limitation, thereby increasing the sensitivity of these hybrid sensors [61]. One example of the

Hybrid WGM methods have been analyzed in recent studies by coupling a WGM resonator to metal nanoparticles [40-42]. One study observed the effects of adding a gold nanoparticle to the equator of a microparticle. The motivation for this hybrid resonator was the need to rapidly detect pathogens. It works based on the principle of creating a plasmonic effect near the equator of the sensor, which enhances the already high-sensing capabilities of WGM-based sensors [41]. Other studies have used triangular gold nanoprisms coupled with WGM sensors. In one case, a gold triangular nanoprism was placed inside a microtoroid WGM resonator. It was shown that the tips of the nanoprism had regions of great plasmonic enhancement. This type of plasmonic enhancement permits the detection of larger protein molecules with high precision as shown in Figure 5 [42].

application of hybrid WGM biosensors is in the determination of proteins. More specifically, it was used to quantify the amount of bovine serum albumin that is absorbed by the gold nanoparticles [62], making the hybrid sensor a perfect candidate for combining plasmonic and high-sensitivity resonant microstructures. In addition, Huckabay et al [63] used WGM resonators to analyze a biomarker for ovarian cancer (CA-125) in a buffer.

Some Other Relevant Examples of SPR/TFBG Applications in the Biomedical Field

SPR is one of the prominent methods used for biomedical applications. Sharma [68] used a sensor based on SPR to detect the concentration of hemoglobin in human blood. Hemoglobin detection is an important medical procedure that has an impact on several clinical methods. Overall, this method of analyzing blood using SPR will lead to its use in blood analysis. Luo et al [69] used a plasmonic method employing gold nanoparticles, improving detection of tumor-targeted cells during X-ray radiotherapy (Figure 6).

Figure 6. Schematic illustration of targeted prostate cancer radiotherapy using prostate-specific membrane antigen (PSMA)-targeted gold nanoparticles (AuNPs) of various sizes. With permission from the Royal Society of Chemistry, 2019 [69].

Others have used gold nanorods to detect breast cancer biomarkers [5,70]. In addition to gold and silver, a few new plasmonic sensors based on different metals, such as magnesium, have been developed recently [71]. TFBG-based sensors have also been used to detect small biomarkers to diagnose lung cancers. The sensor was able to monitor the amplitude shift of sensitive spectrum modes of the TFBG SPR [72]. Furthermore, an immunosensor was used to detect biomarkers for risk stratification and prognosis of heart failure [73]. The detection of drugs and metabolites in patients currently remains a challenge and requires novel tools and methodologies. One study developed a diagnostic system based on silver nanoshells to detect metabolites in biofluids and identify whether patients had postoperative brain infection using embedded gold nanoparticles [74]. By contrast, ELISAs have been used to detect disease biomarkers at ultra-low concentrations. One study used this technique to detect HIV-2 capsid antigen p24 and prostate-specific antigen. This type of cost-effective technique can assist developing countries that require better methods to detect HIV infections. Therefore, it was noted that the plasmonic ELISA is a versatile method of detection for application in biomedical fields [19,75]. Silver nanocubes have also been applied for detecting lung cancer biomarkers, such as microRNAs. microRNAs, which are known to act as tumor

suppressors, can be used for biomedical diagnosis. Zhang et al [49] developed a plasmonic nanoprobe technique to rapidly detect miR-21 biomarkers. miR-21 was used as a biomarker for diagnosing lung cancer early. The technique was based on the LSRP spectral shift that was caused by a change in the refractive index. Plasmon-based sensors are very versatile and in the near future it will be possible to see robust and cheap point-of-care devices for various daily monitoring and diagnosis of different medical conditions [19].

Conclusions and Future Prospects

In this brief review paper, different plasmonic sensing methods and biosensing applications were discussed. Overall, biosensing is an attractive research area and novel sensing methods are being developed rapidly. Biosensing is a very powerful technique and will have a substantial impact on the biomedical community. This review summarized current methods and results that have influenced applications based on plasmonic biosensors. It was observed that the SPR is a notable principle for biosensing. This method is used for different applications such as for detection of sweat loss, biomarkers, and even hemoglobin concentration in human blood. Plasmon-based biosensors are versatile and will continue to be investigated and developed with technological advancements in the future to improve selectivity and robustness.

Conflicts of Interest

None declared.

References

1. Virk M, Xiong K, Svedendahl M, Käll M, Dahlin AB. A thermal plasmonic sensor platform: resistive heating of nanohole arrays. *Nano Lett* 2014 Jun 11;14(6):3544-3549. [doi: [10.1021/nl501154z](https://doi.org/10.1021/nl501154z)] [Medline: [24807397](https://pubmed.ncbi.nlm.nih.gov/24807397/)]

2. Ma Y, Farrell G, Semenova Y, Wu Q. Analysis and applications of nanocavity structures used as tunable filters and sensors. *Infrared Physics & Technology* 2012 Sep;55(5):389-394. [doi: [10.1016/j.infrared.2012.07.005](https://doi.org/10.1016/j.infrared.2012.07.005)]
3. Nouri-Novin S, Sadatgol M, Zarrabi FB, Bazgir M. A hollow rectangular plasmonic absorber for nano biosensing applications. *Optik* 2019 Jan;176:14-23. [doi: [10.1016/j.jpleo.2018.09.065](https://doi.org/10.1016/j.jpleo.2018.09.065)]
4. Tong L, Wei H, Zhang S, Xu H. Recent advances in plasmonic sensors. *Sensors (Basel)* 2014 May 05;14(5):7959-7973 [FREE Full text] [doi: [10.3390/s140507959](https://doi.org/10.3390/s140507959)] [Medline: [24803189](https://pubmed.ncbi.nlm.nih.gov/24803189/)]
5. Sugumaran S, Jamlos M, Ahmad M, Bellan C, Schreurs D. Nanostructured materials with plasmonic nanobiosensors for early cancer detection: A past and future prospect. *Biosens Bioelectron* 2018 Feb 15;100:361-373. [doi: [10.1016/j.bios.2017.08.044](https://doi.org/10.1016/j.bios.2017.08.044)] [Medline: [28946108](https://pubmed.ncbi.nlm.nih.gov/28946108/)]
6. Zeng X, Yang Y, Zhang N, Ji D, Gu X, Jornet JM, et al. Plasmonic Interferometer Array Biochip as a New Mobile Medical Device for Cancer Detection. *IEEE J. Select. Topics Quantum Electron* 2019 Jan;25(1):1-7. [doi: [10.1109/jstqe.2018.2865418](https://doi.org/10.1109/jstqe.2018.2865418)]
7. Zhao F, Zeng J, Shih W. Nanoporous Gold Nanocomposites as a Versatile Platform for Plasmonic Engineering and Sensing. *Sensors (Basel)* 2017 Jun 28;17(7):1519 [FREE Full text] [doi: [10.3390/s17071519](https://doi.org/10.3390/s17071519)] [Medline: [28657586](https://pubmed.ncbi.nlm.nih.gov/28657586/)]
8. Xavier J, Vincent S, Meder F, Vollmer F. Advances in optoplasmonic sensors-combining optical nano/microcavities and photonic crystals with plasmonic nanostructures and nanoparticles. *Nanophotonics* 2018;7(1):1-38 [FREE Full text] [doi: [10.1515/nanoph-2017-0064](https://doi.org/10.1515/nanoph-2017-0064)]
9. Liu N, Mesch M, Weiss T, Hentschel M, Giessen H. Infrared perfect absorber and its application as plasmonic sensor. *Nano Lett* 2010 Jul 14;10(7):2342-2348. [doi: [10.1021/nl9041033](https://doi.org/10.1021/nl9041033)] [Medline: [20560590](https://pubmed.ncbi.nlm.nih.gov/20560590/)]
10. Carnovale C, Bryant G, Shukla R, Bansal V. Identifying Trends in Gold Nanoparticle Toxicity and Uptake: Size, Shape, Capping Ligand, and Biological Corona. *ACS Omega* 2019 Jan 04;4(1):242-256. [doi: [10.1021/acsomega.8b03227](https://doi.org/10.1021/acsomega.8b03227)]
11. Ferdous Z, Nemmar A. Health Impact of Silver Nanoparticles: A Review of the Biodistribution and Toxicity Following Various Routes of Exposure. *Int J Mol Sci* 2020 Mar 30;21(7):2375 [FREE Full text] [doi: [10.3390/ijms21072375](https://doi.org/10.3390/ijms21072375)] [Medline: [32235542](https://pubmed.ncbi.nlm.nih.gov/32235542/)]
12. Lee D, Jeong SH, Kang E. Nanodiamond/gold nanorod nanocomposites with tunable light-absorptive and local plasmonic properties. *Journal of Industrial and Engineering Chemistry* 2018 Sep;65:205-212. [doi: [10.1016/j.jiec.2018.04.030](https://doi.org/10.1016/j.jiec.2018.04.030)]
13. Tokel O, Yildiz UH, Inci F, Durmus NG, Ekiz OO, Turker B, et al. Portable microfluidic integrated plasmonic platform for pathogen detection. *Sci Rep* 2015 Mar 24;5(1):9152 [FREE Full text] [doi: [10.1038/srep09152](https://doi.org/10.1038/srep09152)] [Medline: [25801042](https://pubmed.ncbi.nlm.nih.gov/25801042/)]
14. Cetin A, Coskun A, Galarreta B, Huang M, Herman D, Ozcan A, et al. Handheld high-throughput plasmonic biosensor using computational on-chip imaging. *Light Sci Appl* 2014 Jan 3;3(1):e122-e122. [doi: [10.1038/lsa.2014.3](https://doi.org/10.1038/lsa.2014.3)]
15. Hong L, Li H, Yang H, Sengupta K. Fully Integrated Fluorescence Biosensors On-Chip Employing Multi-Functional Nanoplasmonic Optical Structures in CMOS. *IEEE J. Solid-State Circuits* 2017 Sep;52(9):2388-2406. [doi: [10.1109/jssc.2017.2712612](https://doi.org/10.1109/jssc.2017.2712612)]
16. Katzer C, Grosse V, Schmidl F, Michalowski P, Schmidl G, Mueller R, et al. YBa₂Cu₃O₇- δ matrix-induced in situ growth of plasmonic Au nanoparticles for biological sensor devices. *J Nanopart Res* 2012 Nov 10;14(12):1285. [doi: [10.1007/s11051-012-1285-7](https://doi.org/10.1007/s11051-012-1285-7)]
17. Anker JN, Hall WP, Lyandres O, Shah NC, Zhao J, Van Duyne RP. Biosensing with plasmonic nanosensors. *Nat Mater* 2008 Jun;7(6):442-453. [doi: [10.1038/nmat2162](https://doi.org/10.1038/nmat2162)] [Medline: [18497851](https://pubmed.ncbi.nlm.nih.gov/18497851/)]
18. Rodrigues D, Barbosa AI, Rebelo R, Kwon IK, Reis RL, Correlo VM. Skin-Integrated Wearable Systems and Implantable Biosensors: A Comprehensive Review. *Biosensors (Basel)* 2020 Jul 21;10(7):79 [FREE Full text] [doi: [10.3390/bios10070079](https://doi.org/10.3390/bios10070079)] [Medline: [32708103](https://pubmed.ncbi.nlm.nih.gov/32708103/)]
19. Tokel O, Inci F, Demirci U. Advances in plasmonic technologies for point of care applications. *Chem Rev* 2014 Jun 11;114(11):5728-5752 [FREE Full text] [doi: [10.1021/cr4000623](https://doi.org/10.1021/cr4000623)] [Medline: [24745365](https://pubmed.ncbi.nlm.nih.gov/24745365/)]
20. Sharma A, Mohr G. Plasmonic optical sensor for determination of refractive index of human skin tissues. *Sensors and Actuators B: Chemical* 2016 Apr;226:312-317. [doi: [10.1016/j.snb.2015.11.119](https://doi.org/10.1016/j.snb.2015.11.119)]
21. Rifat A, Ahmed R, Yetisen A, Butt H, Sabouri A, Mahdiraji G, et al. Photonic crystal fiber based plasmonic sensors. *Sensors and Actuators B: Chemical* 2017 May;243:311-325. [doi: [10.1016/j.snb.2016.11.113](https://doi.org/10.1016/j.snb.2016.11.113)]
22. Sharma A, Gupta A. Design of a plasmonic optical sensor probe for humidity-monitoring. *Sensors and Actuators B: Chemical* 2013 Nov;188:867-871. [doi: [10.1016/j.snb.2013.08.002](https://doi.org/10.1016/j.snb.2013.08.002)]
23. Shevchenko Y, Camci-Unal G, Cuttica DF, Dokmeci MR, Albert J, Khademhosseini A. Surface plasmon resonance fiber sensor for real-time and label-free monitoring of cellular behavior. *Biosens Bioelectron* 2014 Jun 15;56:359-367 [FREE Full text] [doi: [10.1016/j.bios.2014.01.018](https://doi.org/10.1016/j.bios.2014.01.018)] [Medline: [24549115](https://pubmed.ncbi.nlm.nih.gov/24549115/)]
24. Sharma AK, Pandey AK, Kaur B. A Review of advancements (2007–2017) in plasmonics-based optical fiber sensors. *Optical Fiber Technology* 2018 Jul;43:20-34. [doi: [10.1016/j.yofte.2018.03.008](https://doi.org/10.1016/j.yofte.2018.03.008)]
25. Popescu VA, Puscas NN, Perrone G. Simulation of the Sensing Performance of a Plasmonic Biosensor Based on Birefringent Solid-Core Microstructured Optical Fiber. *Plasmonics* 2016 Aug 8;12(3):905-911. [doi: [10.1007/s11468-016-0342-y](https://doi.org/10.1007/s11468-016-0342-y)]
26. Popescu VA. Application of a Plasmonic Biosensor for Detection of Human Blood Groups. *Plasmonics* 2016 Nov 22;12(6):1733-1739. [doi: [10.1007/s11468-016-0440-x](https://doi.org/10.1007/s11468-016-0440-x)]
27. Momota MR, Hasan MR. Hollow-core silver coated photonic crystal fiber plasmonic sensor. *Optical Materials* 2018 Feb;76:287-294. [doi: [10.1016/j.optmat.2017.12.049](https://doi.org/10.1016/j.optmat.2017.12.049)]

28. Derbali J, AbdelMalek F, Obayya SSA, Bouchriha H, Letizia R. Design of a compact photonic crystal sensor. *Opt Quant Electron* 2010 Dec 24;42(8):463-472. [doi: [10.1007/s11082-010-9429-4](https://doi.org/10.1007/s11082-010-9429-4)]
29. Lim WQ, Gao Z. Plasmonic nanoparticles in biomedicine. *Nano Today* 2016 Apr;11(2):168-188. [doi: [10.1016/j.nantod.2016.02.002](https://doi.org/10.1016/j.nantod.2016.02.002)]
30. Xu H, Kan C, Wei J, Ni Y, Miao C, Wang C, et al. Synthesis and Plasmonic Property of Ag Nanorods. *Plasmonics* 2016 Jun 2;11(6):1645-1652. [doi: [10.1007/s11468-016-0257-7](https://doi.org/10.1007/s11468-016-0257-7)]
31. Lim J, Majetich S. Composite magnetic-plasmonic nanoparticles for biomedicine: Manipulation and imaging. *Nano Today* 2013 Feb;8(1):98-113. [doi: [10.1016/j.nantod.2012.12.010](https://doi.org/10.1016/j.nantod.2012.12.010)]
32. Yockell-Lelièvre H, Bukar N, McKeating KS, Arnaud M, Cosin P, Guo Y, et al. Plasmonic sensors for the competitive detection of testosterone. *Analyst* 2015 Aug 07;140(15):5105-5111. [doi: [10.1039/c5an00694e](https://doi.org/10.1039/c5an00694e)] [Medline: [26034786](https://pubmed.ncbi.nlm.nih.gov/26034786/)]
33. Kamada S, Okamoto T, El-Zohary SE, Mori A, Haraguchi M. Design optimization and resonance modes of a plasmonic sensor based on a rectangular resonator. *Optics Communications* 2018 Nov;427:220-225. [doi: [10.1016/j.optcom.2018.06.076](https://doi.org/10.1016/j.optcom.2018.06.076)]
34. Spillane SM, Kippenberg TJ, Painter OJ, Vahala KJ. Ideality in a Fiber-Taper-Coupled Microresonator System for Application to Cavity Quantum Electrodynamics. *Phys. Rev. Lett* 2003 Jul 22;91(4):043902. [doi: [10.1103/physrevlett.91.043902](https://doi.org/10.1103/physrevlett.91.043902)]
35. Cai M, Painter O, Vahala KJ. Observation of Critical Coupling in a Fiber Taper to a Silica-Microsphere Whispering-Gallery Mode System. *Phys. Rev. Lett* 2000 Jul 3;85(1):74-77. [doi: [10.1103/physrevlett.85.74](https://doi.org/10.1103/physrevlett.85.74)]
36. Lee M, Jeon H, Kim S. A highly tunable and fully biocompatible silk nanoplasmonic optical sensor. *Nano Lett* 2015 May 13;15(5):3358-3363. [doi: [10.1021/acs.nanolett.5b00680](https://doi.org/10.1021/acs.nanolett.5b00680)] [Medline: [25821994](https://pubmed.ncbi.nlm.nih.gov/25821994/)]
37. Ding L, Gao Y, Di J. A sensitive plasmonic copper(II) sensor based on gold nanoparticles deposited on ITO glass substrate. *Biosens Bioelectron* 2016 Sep 15;83:9-14. [doi: [10.1016/j.bios.2016.04.002](https://doi.org/10.1016/j.bios.2016.04.002)] [Medline: [27093484](https://pubmed.ncbi.nlm.nih.gov/27093484/)]
38. Tu M, Sun T, Grattan K. LSPR optical fibre sensors based on hollow gold nanostructures. *Sensors and Actuators B: Chemical* 2014 Feb;191:37-44. [doi: [10.1016/j.snb.2013.09.094](https://doi.org/10.1016/j.snb.2013.09.094)]
39. Yousuf SMEH, Sakib MA, Islam MZ. A High-Performance Plasmonic Nanosensor Based on an Elliptical Nanorod in an MIM Configuration. *IEEE Sensors J* 2018 Aug 1;18(15):6145-6153. [doi: [10.1109/jsen.2018.2847760](https://doi.org/10.1109/jsen.2018.2847760)]
40. Manzo M, Schwend R. A Novel Microlaser-Based Plasmonic-Polymer Hybrid Resonator for Multiplexed Biosensing Applications. *ASME Journal of Medical Diagnostics* 2019 May;2(2):021006. [doi: [10.1115/1.4042377](https://doi.org/10.1115/1.4042377)]
41. Shopova SI, Rajmangal R, Holler S, Arnold S. Plasmonic enhancement of a whispering-gallery-mode biosensor for single nanoparticle detection. *Appl. Phys. Lett* 2011 Jun 13;98(24):243104. [doi: [10.1063/1.3599584](https://doi.org/10.1063/1.3599584)]
42. Nadgaran H, Afkhami Garaei M. Enhancement of a whispering gallery mode microtoroid resonator by plasmonic triangular gold nanoprism for label-free biosensor applications. *Journal of Applied Physics* 2015 Jul 28;118(4):043101. [doi: [10.1063/1.4927266](https://doi.org/10.1063/1.4927266)]
43. Chung T, Lee S, Song E, Chun H, Lee B. Plasmonic nanostructures for nano-scale bio-sensing. *Sensors (Basel)* 2011;11(11):10907-10929 [FREE Full text] [doi: [10.3390/s111110907](https://doi.org/10.3390/s111110907)] [Medline: [22346679](https://pubmed.ncbi.nlm.nih.gov/22346679/)]
44. Liu Y, Park Y, Lee SE. Thermo-responsive mechano-optical plasmonic nano-antenna. *Appl. Phys. Lett* 2016 Jul 04;109(1):013109. [doi: [10.1063/1.4954907](https://doi.org/10.1063/1.4954907)]
45. Yuan H, Register JK, Wang H, Fales AM, Liu Y, Vo-Dinh T. Plasmonic nanoprobe for intracellular sensing and imaging. *Anal Bioanal Chem* 2013 Jul 11;405(19):6165-6180. [doi: [10.1007/s00216-013-6975-1](https://doi.org/10.1007/s00216-013-6975-1)] [Medline: [23665636](https://pubmed.ncbi.nlm.nih.gov/23665636/)]
46. Schopf C, Martín A, Iacopino D. Plasmonic detection of mercury via amalgam formation on surface-immobilized single Au nanorods. *Sci Technol Adv Mater* 2017 Jan 09;18(1):60-67 [FREE Full text] [doi: [10.1080/14686996.2016.1258293](https://doi.org/10.1080/14686996.2016.1258293)] [Medline: [28179959](https://pubmed.ncbi.nlm.nih.gov/28179959/)]
47. Tao H, Lin Y, Yan J, Di J. A plasmonic mercury sensor based on silver-gold alloy nanoparticles electrodeposited on indium tin oxide glass. *Electrochemistry Communications* 2014 Mar;40:75-79. [doi: [10.1016/j.elecom.2014.01.002](https://doi.org/10.1016/j.elecom.2014.01.002)]
48. Unser S, Holcomb S, Cary R, Sagle L. Collagen-Gold Nanoparticle Conjugates for Versatile Biosensing. *Sensors (Basel)* 2017 Feb 15;17(2):378 [FREE Full text] [doi: [10.3390/s17020378](https://doi.org/10.3390/s17020378)] [Medline: [28212282](https://pubmed.ncbi.nlm.nih.gov/28212282/)]
49. Zhang L, Zhang Y, Hu Y, Fan Q, Yang W, Li A, et al. Refractive index dependent real-time plasmonic nanoprobe on a single silver nanocube for ultrasensitive detection of the lung cancer-associated miRNAs. *Chem Commun (Camb)* 2015;51(2):294-297. [doi: [10.1039/c4cc06663d](https://doi.org/10.1039/c4cc06663d)] [Medline: [25407574](https://pubmed.ncbi.nlm.nih.gov/25407574/)]
50. Manzo M, Ioppolo T. Untethered photonic sensor for wall pressure measurement. *Opt. Lett* 2015 May 06;40(10):2257. [doi: [10.1364/ol.40.002257](https://doi.org/10.1364/ol.40.002257)]
51. Manzo M. Untethered wall pressure and temperature sensor based on dye doped micro-scale resonator (Southern Methodist University). Ann Arbor, MI: ProQuest Dissertations Publishing; 2015.
52. Ioppolo T, Manzo M. Dome-shaped whispering gallery mode laser for remote wall temperature sensing. *Appl. Opt* 2014 Jul 31;53(22):5065. [doi: [10.1364/ao.53.005065](https://doi.org/10.1364/ao.53.005065)]
53. Manzo M, Cavazos O. Neurotransducers Based Voltage Sensitive Dye-Doped Microlasers. Washington, DC: OSA Publication; 2019 Presented at: Biophotonics Congress: Optics in the Life Sciences Congress 2019 (BODA,BRAIN,NTM,OMA,OMP); 15–17 April 2019; Tucson, AZ. [doi: [10.1364/BODA.2019.JT4A.14](https://doi.org/10.1364/BODA.2019.JT4A.14)]
54. Manzo M, Cavazos O, Ramirez-Cedillo E, Siller H. Embedded Spherical Microlasers for In-vivo Diagnostic Biomechanical Performances. *ASME Journal of Medical Diagnostics* 2020 Nov;3(4):044504. [doi: [10.1115/1.4048466](https://doi.org/10.1115/1.4048466)]

55. Vollmer F, Braun D, Libchaber A, Khoshshima M, Teraoka I, Arnold S. Protein detection by optical shift of a resonant microcavity. *Appl. Phys. Lett* 2002 May 27;80(21):4057-4059. [doi: [10.1063/1.1482797](https://doi.org/10.1063/1.1482797)]
56. Manzo M, Ioppolo T, Ayaz UK, Lapenna V, Ötügen MV. A photonic wall pressure sensor for fluid mechanics applications. *Rev Sci Instrum* 2012 Oct;83(10):105003. [doi: [10.1063/1.4757569](https://doi.org/10.1063/1.4757569)] [Medline: [23126796](https://pubmed.ncbi.nlm.nih.gov/23126796/)]
57. Wagner HP, Schmitzer H, Lutti J, Borri P, Langbein W. Effects of uniaxial pressure on polar whispering gallery modes in microspheres. *Journal of Applied Physics* 2013 Jun 28;113(24):243101. [doi: [10.1063/1.4811447](https://doi.org/10.1063/1.4811447)]
58. Hanumegowda NM, Stica CJ, Patel BC, White I, Fan X. Refractometric sensors based on microsphere resonators. *Appl. Phys. Lett* 2005 Nov 14;87(20):201107. [doi: [10.1063/1.2132076](https://doi.org/10.1063/1.2132076)]
59. Li B, Wang Q, Xiao Y, Jiang X, Li Y, Xiao L, et al. On chip, high-sensitivity thermal sensor based on high-Q polydimethylsiloxane-coated microresonator. *Appl. Phys. Lett* 2010 Jun 21;96(25):251109. [doi: [10.1063/1.3457444](https://doi.org/10.1063/1.3457444)]
60. Fan H, Xia C, Fan L, Wang L, Shen M. Graphene-supported plasmonic whispering-gallery mode in a metal-coated microcavity for sensing application with ultrahigh sensitivity. *Optics Communications* 2018 Mar;410:668-673. [doi: [10.1016/j.optcom.2017.11.018](https://doi.org/10.1016/j.optcom.2017.11.018)]
61. Foreman MR, Vollmer F. Theory of resonance shifts of whispering gallery modes by arbitrary plasmonic nanoparticles. *New J. Phys* 2013 Aug 02;15(8):083006. [doi: [10.1088/1367-2630/15/8/083006](https://doi.org/10.1088/1367-2630/15/8/083006)]
62. Santiago-Cordoba MA, Boriskina SV, Vollmer F, Demirel MC. Nanoparticle-based protein detection by optical shift of a resonant microcavity. *Appl. Phys. Lett* 2011 Aug 15;99(7):073701. [doi: [10.1063/1.3599706](https://doi.org/10.1063/1.3599706)]
63. Huckabay HA, Wildgen SM, Dunn RC. Label-free detection of ovarian cancer biomarkers using whispering gallery mode imaging. *Biosens Bioelectron* 2013 Jul 15;45:223-229. [doi: [10.1016/j.bios.2013.01.072](https://doi.org/10.1016/j.bios.2013.01.072)] [Medline: [23500368](https://pubmed.ncbi.nlm.nih.gov/23500368/)]
64. Chen X, Nan Y, Ma X, Liu H, Liu W, Shi L, et al. Detection of Small Biomolecule Interactions Using a Plasmonic Tilted Fiber Grating Sensor. *J. Lightwave Technol* 2019 Jun 1;37(11):2792-2799. [doi: [10.1109/jlt.2018.2870337](https://doi.org/10.1109/jlt.2018.2870337)]
65. Guo T, Liu F, Liang X, Qiu X, Huang Y, Xie C, et al. Highly sensitive detection of urinary protein variations using tilted fiber grating sensors with plasmonic nanocoatings. *Biosens Bioelectron* 2016 Apr 15;78:221-228 [FREE Full text] [doi: [10.1016/j.bios.2015.11.047](https://doi.org/10.1016/j.bios.2015.11.047)] [Medline: [26618641](https://pubmed.ncbi.nlm.nih.gov/26618641/)]
66. Zhang Y, Wang F, Qian S, Liu Z, Wang Q, Gu Y, et al. A Novel Fiber Optic Surface Plasmon Resonance Biosensors with Special Boronic Acid Derivative to Detect Glycoprotein. *Sensors (Basel)* 2017 Oct 01;17(10):2259 [FREE Full text] [doi: [10.3390/s17102259](https://doi.org/10.3390/s17102259)] [Medline: [28974028](https://pubmed.ncbi.nlm.nih.gov/28974028/)]
67. Dong C, He L, Xiao Y, Gaddam VR, Ozdemir SK, Han Z, et al. Fabrication of high-Q polydimethylsiloxane optical microspheres for thermal sensing. *Appl. Phys. Lett* 2009 Jun 08;94(23):231119. [doi: [10.1063/1.3152791](https://doi.org/10.1063/1.3152791)]
68. Sharma A. Plasmonic biosensor for detection of hemoglobin concentration in human blood: Design considerations. *Journal of Applied Physics* 2013 Jul 28;114(4):044701. [doi: [10.1063/1.4816272](https://doi.org/10.1063/1.4816272)]
69. Luo D, Wang X, Zeng S, Ramamurthy G, Burda C, Basilion J. Prostate-specific membrane antigen targeted gold nanoparticles for prostate cancer radiotherapy: does size matter for targeted particles? *Chem. Sci* 2019 Sep 11;10(35):8119-8128. [doi: [10.1039/C9SC02290B](https://doi.org/10.1039/C9SC02290B)]
70. Chen S, Zhao Q, Zhang L, Wang L, Zeng Y, Huang H. Combined detection of breast cancer biomarkers based on plasmonic sensor of gold nanorods. *Sensors and Actuators B: Chemical* 2015 Dec;221:1391-1397. [doi: [10.1016/j.snb.2015.08.023](https://doi.org/10.1016/j.snb.2015.08.023)]
71. Li R, Xie S, Zhang L, Li L, Kong D, Wang Q, et al. Soft and transient magnesium plasmonics for environmental and biomedical sensing. *Nano Res* 2018 Mar 21;11(8):4390-4400. [doi: [10.1007/s12274-018-2028-6](https://doi.org/10.1007/s12274-018-2028-6)]
72. Ribaut C, Voisin V, Malachovská V, Dubois V, Mégret P, Wattiez R, et al. Small biomolecule immunosensing with plasmonic optical fiber grating sensor. *Biosens Bioelectron* 2016 Mar 15;77:315-322. [doi: [10.1016/j.bios.2015.09.019](https://doi.org/10.1016/j.bios.2015.09.019)] [Medline: [26432194](https://pubmed.ncbi.nlm.nih.gov/26432194/)]
73. Luo B, Wu S, Zhang Z, Zou W, Shi S, Zhao M, et al. Human heart failure biomarker immunosensor based on excessively tilted fiber gratings. *Biomed. Opt. Express* 2016 Dec 05;8(1):57. [doi: [10.1364/boe.8.000057](https://doi.org/10.1364/boe.8.000057)]
74. Kuruvinashetti K, Kashani AS, Badilescu S, Beaudet D, Piekny A, Packirisamy M. Intracellular Localized Surface Plasmonic Sensing for Subcellular Diagnosis. *Plasmonics* 2017 Dec 18;13(5):1639-1648. [doi: [10.1007/s11468-017-0673-3](https://doi.org/10.1007/s11468-017-0673-3)]
75. de la Rica R, Stevens MM. Plasmonic ELISA for the ultrasensitive detection of disease biomarkers with the naked eye. *Nat Nanotechnol* 2012 Dec 28;7(12):821-824. [doi: [10.1038/nnano.2012.186](https://doi.org/10.1038/nnano.2012.186)] [Medline: [23103935](https://pubmed.ncbi.nlm.nih.gov/23103935/)]

Abbreviations

AdoHcy: S-adenosyl-L-homocysteine
ATR: attenuated total reflection
LSPR: localized surface plasmon resonance
SPR: surface plasmon resonance
TFBG: tilted fiber Bragg grating
WGM: whispering gallery mode

Edited by G Eysenbach; submitted 13.01.20; peer-reviewed by A Civit, A Aminbeidokhti; comments to author 10.03.20; revised version received 01.10.20; accepted 11.03.21; published 12.04.21

Please cite as:

Manzo M, Cavazos O, Huang Z, Cai L

Plasmonic and Hybrid Whispering Gallery Mode-Based Biosensors: Literature Review

JMIR Biomed Eng 2021;6(2):e17781

URL: <https://biomedeng.jmir.org/2021/2/e17781>

doi: [10.2196/17781](https://doi.org/10.2196/17781)

PMID:

©Maurizio Manzo, Omar Cavazos, Zhenhua Huang, Liping Cai. Originally published in JMIR Biomedical Engineering (<http://biomedeng.jmir.org>), 12.04.2021. This is an open-access article distributed under the terms of the Creative Commons Attribution License (<https://creativecommons.org/licenses/by/4.0/>), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work, first published in JMIR Biomedical Engineering, is properly cited. The complete bibliographic information, a link to the original publication on <http://biomedeng.jmir.org/>, as well as this copyright and license information must be included.